


Parallax Laser, Sept. 11, 2003



- Power including all = 24.5 ± 0.38 . Variation = 3.8%
- Power after red line = 24.7 ± 0.1 , Variation = 0.8%

Tests of a free running PLX50 co2 laser by Dr. Anastassiou of OmniGuide Communications (Cambridge, US) shows power stability of better than $\pm 0.4\%$ after 6 minutes warm up time.

----- Courtesy of OmniGuide Communications